

校准证书
CALIBRATION CERTIFICATE

证书编号:

Certificate No.



J202512241371-0001

第 1 页 共 4 页

Page

of

委托方

Client

ABB Electrical Industries (ABB Arab) S.A.E.

联络信息

Contact Inf.

Building 208, North 90 Street, 5th Settlement, New Cairo, Egypt

仪器名称

Description

筛网(试验筛)

型号/规格

Model/Type

200目/0.074mm

制造厂

Manufacturer

出厂编号

Serial No.

管理号

Asset No.

529892

接收日期

Receipt Date

2025年12月24日

Y M D

校准日期

Cal. Date

2025年12月25日

Y M D

发布日期

Issued Date

2025年12月25日

Y M D

批准

Approved by

赖志文

赖志文

审核

Inspected by

伍水生

伍水生

校准

Calibrated by

陈泽康

陈泽康

证书专用章

(Stamp)

总部地址(Headquarters Add.): 广东省广州市番禺区创运路8号

No.8.Chuangyun Rd,Panyu District,Guangzhou,Guangdong,China

实验室地址(Add.of the Lab): 广东省广州市番禺区创运路8号

No.8 Chuangyun Rd,Panyu District,Guangzhou,Guangdong,China

联系电话(Tel.):400-602-0999

邮政编码(Postcode):511450

网站(Website):http:// www.grgtest.com

电子邮件(E-mail):grgtest@grgtest.com



扫一扫验真伪

校验码: 590521

校准说明 DIRECTIONS OF CALIBRATION

证书编号: J202512241371-0001

第 2 页 共 4 页

Certificate No.

Page of

- 1.本实验室的质量管理体系符合ISO/IEC 17025:2017标准的要求,校准结果均可溯源至国际单位制(SI)单位。
(The quality system is in accordance with ISO/IEC 17025:2017,the calibration results are traceable to the International System of Units (SI).)
 - 2.本结果仅对本次校准样品有效。未经实验室批准,不得部分复制。如有疑问请在15个工作日内反馈。
(The result is only valid for the calibrated sample.The certificate shall not be reproduced except in full,without the written approval of our laboratory .please feedback to us within 15 days if you have any question.)
 - 3.本证书编号具有唯一性,后缀若带有“-Gx”的证书为替换证书,自发出后原证书即刻作废,修改后的证书以客户端内容为准。(Each certificate has a unique number. The suffix of "-Gx" will be added to the number as a replacement of the old version. The original certificate will be officially invalid once the new certificate number is issued.The modified certificate shall be based on the client content.)
 - 4.证书中最大允许误差、判定结果仅供参考,其中“P”代表“合格”,“F”代表“不合格”,“N/A”代表“不适用”。使用人员应结合实际测量需求,评估测量不确定度对符合性评定的影响。(MPE & judgement result in the datasheet is only for reference , "P" is "Pass" , "F" is "Fail" and "N/A" is "Not Applicable".Whereas users should evaluate the effects of MU of calibration results on conformance assessment by actual measurement.)
 - 5.校准地点、环境条件(Place and environmental conditions of the calibration):
地点: 广州计量精密测量室
Place Guangzhou Metrology Precision Measurement Room
温度: 20.3℃ 相对湿度: 51%
Temperature Relative Humidity
 - 6.建议复校时间间隔: 1年,送校单位也可按实际使用情况自主决定。
Suggested calibration interval is 1 year or it can be altered depending on the actual usage of the user.
 - 7.本次校准的技术依据及CNAS认可范围,超出范围的内容未被认可。详细认可范围请查看CNAS网站证书附件。(Reference document and accredited scope by CNAS for calibration, beyond which isn't accredited. Please see the attachment of certificate on CNAS website for details.)
- JJF 1175-2021 试验筛校准规范(C.S. for Test Sieves) 长度: (0.04~125)mm

校准说明
DIRECTIONS OF CALIBRATION

证书编号: J202512241371-0001

第 3 页 共 4 页

Certificate No.

Page of

8. 本次校准使用的主要测量标准(Main Standards of Measurement Used in the Calibration.):

名称 Description	编号 Serial No.	证书号/有效期 Certificate No./ Due Date	溯源机构 Traceability Institute	技术特征 Technique Character
图像处理万能工具显微镜 Toolmaker's Microscope	BZ0008	CYY202500434 2026-05-25	广东省计量科学 研究院/SCM	(200×100) mm MPE:± (1+L/100) μm

校准结果 RESULTS OF CALIBRATION

证书编号: J202512241371-0001

第 4 页 共 4 页

Certificate No.

Page of

校准项目 Subject	校准结果 Calibration Results	技术指标 Specification	结论 Conclusion
1. 外观及各部分相互作用: Appearance and interreaction:	符合 Pass	无影响计量性能的因素 No factors affecting measurement performance	P

2. 金属丝编织网试验筛校准

Calibration of Woven Wire Test Sieve Cloth

校准项目 Item	实测值(mm) Measured Value	不确定度(mm) $U(k=2)$
筛孔最大尺寸 Max dimension	0.084	0.005
筛孔平均尺寸 Average dimension		
经线方向 Engthwise	0.075	0.005
纬线方向 Crosswise	0.077	0.005
金属丝直径 Average dimension	0.051	0.005

备注:

Notes:

1. 校准结论: 按校准结果使用

Conclusion: Reference the calibration results

2. 1.本报告给出的扩展不确定度是由合成标准不确定度乘以包含概率约为95%时对应的包含因子 k 得到的

(The extended uncertainty given in the report is obtained by multiplying the combined standard uncertainty by the coverage factor k corresponding to the inclusion probability of about 95%)

3. 依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

以下空白(Following blank)