

校准证书

CALIBRATION CERTIFICATE

证书编号:

Certificate No.



J202408090103-0001

第 1 页 共 5 页

Page of

委托方

Client

Danesh Control Farda Research and Lab Company

联络信息

Contact Inf.

No. 20, Sahand Azaran Industrial Town, The beginning os Basmenj Road, Police office of Tehran Road, Tabriz, East Azerbaijan Province, Iran

仪器名称

Description

Impulse Voltage Tester (冲击电压测试仪)

型号/规格

Model/Type

SUG335

制造厂

Manufacturer

LISUNGROUP

出厂编号

Serial No.

LSG3352495

管理号

Asset No.

接收日期

Receipt Date

2024年08月30日

校准日期

Cal. Date

2024年09月02日

Y M D

发布日期

Issued Date

2024年09月02日

Y M D

批准

Approved by

张磊

张磊

审核

Inspected by

焦一鹏

焦一鹏

校准

Calibrated by

彭敬恒

彭敬恒

证书专用章

(Stamp)

总部地址(Headquarters Add.): 广东省广州市黄埔大道西平云路163号

No.163.Pingyun Rd, West of HuangPu Ave.Guangzhou Guangdong China

实验室地址(Add.of the Lab): 江苏无锡太湖国际科技园菱湖大道200号

No.200,Linghu Road,Taihu ISTD,Xinwu District,Wuxi,Jiangsu,China

联系电话(Tel.):400-602-0999

邮政编码(Postcode):510656

网站(Website):http:// www.grgtest.com

电子邮件(E-mail):grgtest@grgtest.com



扫一扫验真伪

校验码: 141004

校准说明 DIRECTIONS OF CALIBRATION

证书编号: J202408090103-0001

第 2 页 共 5 页

Certificate No.

Page of

- 本实验室的质量管理体系符合ISO/IEC 17025:2017标准的要求,校准结果均可溯源至国际单位制(SI)单位。(The quality system is in accordance with ISO/IEC 17025:2017,the calibration results are traceable to the International System of Units (SI).)
 - 本结果仅对本次校准样品有效。未经实验室批准,不得部分复制。如有疑问请在15个工作日内反馈。(The result is only valid for the calibrated sample.The certificate shall not be reproduced except in full,without the written approval of our laboratory .please feedback to us within 15 days if you have any question.)
 - 本证书编号具有唯一性,后缀若带有“-Gx”的证书为替换证书,自发出后原证书即刻作废。(Each certificate has a unique number. The suffix of "-Gx" will be added to the number as a replacement of the old version. The original certificate will be officially invalid once the new certificate number is issued.)
 - 证书中最大允许误差、判定结果仅供参考,其中“P”代表“合格”,“F”代表“不合格”,“N/A”代表“不适用”。使用人员应结合实际测量需求,评估测量不确定度对符合性评定的影响。(MPE & judgement result in the datasheet is only for reference , "P" is "Pass" , "F" is "Fail" and "N/A" is "Not Applicable".Whereas users should evaluate the effects of MU of calibration results on conformance assessment by actual measurement.)
 - 校准地点、环境条件(Place and environmental conditions of the calibration):
地点: 无锡计量电子室
Place Wuxi Metrology Electronics Room
温度: 21℃ 相对湿度: 51%
Temperature Relative Humidity
 - 建议复校时间间隔: 1年,送校单位也可按实际使用情况自主决定。
Suggested calibration interval is 1 year or it can be altered depending on the actual usage of the user.
 - 本次校准的技术依据及CNAS认可范围,超出范围的内容未被认可。详细认可范围请查看CNAS网站证书附件。(Reference document and accredited scope by CNAS for calibration, beyond which isn't accredited. Please see the attachment of certificate on CNAS website for details.)
- IEC 61000-4-5-2014 电磁兼容第4-5部分 试验和测量技术 浪涌(冲击)抗扰度试验(Electromagnetic compatibility(EMC)-Part 4-5:Testing and measurement techniques-Surge immunity test) 开路电压: (0.5 ~40)kV
短路电流: (0.25~50)kA 上升时间: 1ns~5s

校准说明
DIRECTIONS OF CALIBRATION

证书编号: J202408090103-0001

第 3 页 共 5 页

Certificate No.

Page of

8. 本次校准使用的主要测量标准(Main Standards of Measurement Used in the Calibration.):

名称	编号	证书号/有效期	溯源机构	技术特征
Description	Serial No.	Certificate No./ Due Date	Traceability Institute	Technique Character
示波器 Oscilloscope	C023849	J202406051047- 0003 2025-06-04	广电计量检测集 团股份有限公司	V:±1% , t: ±0.01%
高压探头	C066608	J202408086916- 0003 2025-08-16	广电计量检测集 团股份有限公司	MPE: ±3%

9. 计量溯源性声明(Measurement traceability declaration.):

示波器/Oscilloscope(C023849)→多功能校准器(多功能校准源、标准功率源、示波器校准仪)

(6027905)→精密交直流同轴分流器(05155847)→100 kHz 以下交
流电流基准装置(中国计量科学研究院/NIM);

高压探头(C066608)→交直流高压源系统(标准电压互感器)/Standard potential transformer(007300)→交直流高
压源系统(标准电压互感器)/Standard potential transformer(007300);高压探头(C066608)→交直流高压源系统
(直流标准电阻分压器)/DC standard resistor divider(190810)→交直流高压源系统(直流标准电阻分压器)/DC
standard resistor divider(190810);

校准结果
RESULTS OF CALIBRATION

证书编号: J202408090103-0001

第 4 页 共 5 页

Certificate No.

Page of

1、外观以及一般性检查: 正常

In view of External and Generality check : Pass

2、浪涌(Surge)

2.1 开路脉冲电压的校准(Calibration of Open Circuit Voltage):

标称值 Nominal (kV)	实测值 Measured (kV)	误差 Error (kV)	不确定度 $U_{rel}(k=2)$ (%)	允许误差 MPE (kV)	结论 Conclusion (Pass/Fail)
0.5	0.51	-0.01	3.5	± 0.05	P
1.0	1.00	0.00	3.5	± 0.10	P
2.0	2.05	-0.05	3.5	± 0.20	P
4.0	4.08	-0.08	3.5	± 0.40	P
6.0	6.04	-0.04	3.5	± 0.60	P
8.0	8.08	-0.08	3.5	± 0.80	P
10.0	10.16	-0.16	3.5	± 1.00	P
12.0	12.30	-0.30	3.5	± 1.20	P
-0.5	-0.49	-0.01	3.5	± 0.05	P
-1.0	-0.98	-0.02	3.5	± 0.10	P
-2.0	-1.95	-0.05	3.5	± 0.20	P
-4.0	-3.89	-0.11	3.5	± 0.40	P
-6.0	-5.84	-0.16	3.5	± 0.60	P
-8.0	-7.77	-0.23	3.5	± 0.80	P
-10.0	-9.72	-0.28	3.5	± 1.00	P
-12.0	-11.70	-0.30	3.5	± 1.20	P

2.2 开路电压脉冲波前时间的校准(Calibration of Open Circuit Front Time):

电压 Voltage (kV)	标称值 Nominal (μ s)	实测值 Measured (μ s)	误差 Error (μ s)	不确定度 $U_{rel}(k=2)$ (%)	允许误差 MPE (μ s)	结论 Conclusion (Pass/Fail)
0.5	1.2	1.24	-0.04	5.5	± 0.36	P
1.0	1.2	1.18	0.02	5.5	± 0.36	P
2.0	1.2	1.09	0.11	5.5	± 0.36	P
4.0	1.2	1.04	0.16	5.5	± 0.36	P
6.0	1.2	1.06	0.14	5.5	± 0.36	P
8.0	1.2	1.12	0.08	5.5	± 0.36	P
10.0	1.2	1.09	0.11	5.5	± 0.36	P
12.0	1.2	1.10	0.10	5.5	± 0.36	P
-0.5	1.2	1.19	0.01	5.5	± 0.36	P
-1.0	1.2	1.10	0.10	5.5	± 0.36	P
-2.0	1.2	1.06	0.14	5.5	± 0.36	P
-4.0	1.2	1.04	0.16	5.5	± 0.36	P
-6.0	1.2	1.17	0.03	5.5	± 0.36	P
-8.0	1.2	1.08	0.12	5.5	± 0.36	P
10.0	1.2	1.10	0.10	5.5	± 0.36	P
12.0	1.2	1.17	0.03	5.5	± 0.36	P

校准结果
RESULTS OF CALIBRATION

证书编号: J202408090103-0001

第 5 页 共 5 页

Certificate No.

Page of

2.3 开路电压脉冲半波时间的校准(Calibration of Open Circuit Time to Half Value):

电压 Voltage (kV)	标称值 Nominal (μ s)	实测值 Measured (μ s)	误差 Error (μ s)	不确定度 $U_{rel}(k=2)$ (%)	允许误差 MPE (μ s)	结论 Conclusion (Pass/Fail)
0.5	50	58	-8	5.5	± 10	P
1.0	50	57	-7	5.5	± 10	P
2.0	50	57	-7	5.5	± 10	P
4.0	50	58	-8	5.5	± 10	P
6.0	50	59	-9	5.5	± 10	P
8.0	50	58	-8	5.5	± 10	P
10.0	50	58	-8	5.5	± 10	P
12.0	50	57	-7	5.5	± 10	P
-0.5	50	58	-8	5.5	± 10	P
-1.0	50	58	-8	5.5	± 10	P
-2.0	50	59	-9	5.5	± 10	P
-4.0	50	60	-10	5.5	± 10	P
-6.0	50	57	-7	5.5	± 10	P
-8.0	50	60	-10	5.5	± 10	P
-10.0	50	60	-10	5.5	± 10	P
-12.0	50	60	-10	5.5	± 10	P

Notes:

结论(Conclusion): 所校项目符合技术要求

1. 本报告中的扩展不确定度是由标准不确定度乘以包含概率约为95%时的包含因子 k 。The expanded uncertainty is given in the report by the standard uncertainty multiplied by the probability of about 95% when the factor k .

2. 依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

(以下空白)
(The below is blank)