

校准证书
CALIBRATION CERTIFICATE

证书编号:



Certificate No.

J202310101487A-0001

第 1 页 共 4 页

Page

of

委托方

Nemko Scandinavia AS

Client

联络信息

Philip Pedersens vei 11, 1366 Lysaker, Norway

Contact Inf.

仪器名称

Tracking Test Chamber

Description

型号/规格

TTC-1

制造厂

LISUN GROUP

Model/Type

Manufacturer

出厂编号

SOLS20230441-A1

管理号

Serial No.

Asset No.

接收日期

2023年10月19日

校准日期

2023年10月19日

Receipt Date

Y M D

Cal. Date

Y M D

发布日期

2023年10月19日

Issued Date

Y M D

批准

Approved by

李文兴

李文兴

审核

Inspected by

王炳

王炳

校准

Calibrated by

王海彬

王海彬

证书专用章

(Stamp)

总部地址(Headquarters Add.): 广东省广州市黄埔大道西平云路163号

No.163.Pingyun Rd, West of HuangPu Ave.Guangzhou Guangdong China

实验室地址(Add.of the Lab): 广东省广州市黄埔大道西平云路163号

No.163.Pingyun Rd,West of HuangPu Ave.Guangzhou,Guangdong,China

联系电话(Tel.):400-602-0999

邮政编码(Postcode):510656

网站(Website):http:// www.grgtest.com

电子邮件(E-mail):grgtest@grgtest.com



扫一扫验真伪

校准说明 DIRECTIONS OF CALIBRATION

证书编号: J202310101487A-0001

第 2 页 共 4 页

Certificate No.

Page of

- 1.本实验室的质量管理体系符合ISO/IEC 17025:2017标准的要求,校准结果均可溯源至国际单位制(SI)。
(The quality system is in accordance with ISO/IEC 17025:2017,the calibration results are traceable to the International System of Units (SI).)
- 2.本结果仅对本次校准样品有效。未经实验室批准,不得部分复制。如有疑问请在15个工作日内反馈。
(The result is only valid for the calibrated sample.The certificate shall not be reproduced except in full,without the written approval of our laboratroty .please feedback to us within 15 days if you have any question.)
- 3.本证书编号具有唯一性,后缀若带有“-Gx”的证书为替换证书,自发出后原证书即刻作废。
(Each certificate has a unique number. The suffix of "-Gx" will be added to the number as a replacement of the old version. The original certificate will be officially invalid once the new certificate number is issued.)
- 4.证书中最大允许误差、判定结果仅供参考,其中“P”代表“合格”,“F”代表“不合格”,“N/A”代表“不适用”。使用人员应结合实际测量需求,评估测量不确定度对符合性评定的影响。(MPE & judgement result in the datasheet is only for reference, "P" is "Pass", "F" is "Fail" and "N/A" is "Not Applicable".Whereas users should evaluate the effects of MU of calibration results on conformance assessment by actual measurement.)
- 5.本次校准的技术依据及CNAS认可范围,超出范围的内容未被认可。详细认可范围请查看CNAS网站证书附件。(Reference document and accredited scope by CNAS for calibration, beyond which isn't accredited. Please see the attachment of certificate on CNAS website for details.)

JJF(浙)1087-2012 漏电起痕试验仪校准规范(C.S. for Proof Tracking Index Testers) 长度: (1~20)mm 电流: (0.01~1)A; 电压: 0.1V~600V; 角度: (0~90)°; 力值: (0~2)N; 时间: 1s~99s

6. 本次校准使用的主要测量标准(Main Standards of Measurement Used in the Calibration.):

名称	编 号	证书号/有效期	溯源机构	技术特征
Description	Serial No.	Certificate No./ Due Date	Traceability Institute	Technique Character
数字多用表 Digital multimeter	MY57702015	J202212263940- 0023 2024-01-28	广州广电计量检 测股份有限公司	DCV: ±0.0016%;ACV、DCI: ± 0.05%;ACI: ±0.1%;R: ± 0.004%;Freq: ±0.007%
秒表	331766	J202308173772- 0003 2024-08-21	广电计量检测 (无锡)有限公 司	MPE: ±0.5(s/d)
数显卡尺	3107181346	J202307145172- 0029 2024-08-03	广电计量检测集 团股份有限公司	MPE: ±0.03mm
倾角仪	665455 (广电标 识号)	J202308173772- 0057 2024-08-23	广电计量检测集 团股份有限公司	U=0.02° (k=2)

7. 校准地点、环境条件(Place and environmental conditions of the calibration):

地点	客户1楼实验室	温度	23	℃	相对湿度	48	%
Place		Temperature			Relative Humidity		

8. 建议复校时间间隔: 1年, 送校单位也可按实际使用情况自主决定。

Suggested calibration interval is 1 year or it can be altered depending on the actual usage of the user.

校准结果 RESULTS OF CALIBRATION

证书编号: J202310101487A-0001

第 3 页 共 4 页

Certificate No.

Page of

1、外观以及一般性检查: 正常

In view of External and Generality check : Pass

2、试验电压的校准 (Calibration of The Test Voltage):

标称值 Nominal (V)	实测值 Measured (V)	误差 Error (V)	允许误差 MPE (V)	不确定度 $U(k=2)$ (V)	结论 Conclusion (Pass/Fail)
100	100.8	-0.8	± 5.0	0.7	P
200	201.1	-1.1	± 10.0	1.2	P
275	276.8	-1.8	± 13.8	1.3	P
350	350.1	-0.1	± 17.5	1.3	P
425	426.3	-1.3	± 21.3	1.7	P
500	499.1	0.9	± 25.0	1.8	P
600	600	0	± 30	4	P

3、极间电流的校准 (Calibration of Current):

标称值 Nominal (A)	实测值 Measured (A)	误差 Error (A)	允许误差 MPE (A)	不确定度 $U(k=2)$ (A)	结论 Conclusion (Pass/Fail)
0.2	0.200	0.000	± 0.010	0.002	P
0.4	0.400	0.000	± 0.020	0.004	P
0.6	0.598	0.002	± 0.030	0.006	P
0.8	0.798	0.002	± 0.040	0.008	P
1.0	1.000	0.000	± 0.050	0.010	P

4、液滴时间间隔的校准 (Calibration of Time):

标称值 Nominal (s)	实测值 Measured (s)	误差 Error (s)	允许误差 MPE (s)	不确定度 $U(k=2)$ (s)	结论 Conclusion (Pass/Fail)
30	30.16	-0.16	± 1.50	0.26	P

5、电压降:

标称值 Nominal (V)	实测值 Measured (V)	允许误差 MPE (V)	不确定度 $U(k=2)$ (V)	结论 Conclusion (Pass/Fail)
100	3.8	≤ 10.0	1.0	P
200	5.6	≤ 20.0	2.0	P
275	9.2	≤ 27.5	2.8	P
350	10.5	≤ 35.0	3.5	P
425	15.6	≤ 42.5	4.3	P
500	15.2	≤ 50.0	5.0	P
600	17.0	≤ 60.0	6.0	P

校准结果 RESULTS OF CALIBRATION

证书编号: J202310101487A-0001

第 4 页 共 4 页

Certificate No.

Page of

6、电极装置的检查 (Inspection of electrode size)

6.1 电极尺寸的测量

项目	要求	实测值		不确定度 $U(k=2)$
		左电极	右电极	
电极截面尺寸(mm)	$(5 \pm 0.1) \times (2 \pm 0.1)$	5.06 × 2.05	5.05 × 2.01	0.03mm
电极长(mm)	≥ 12	12.5	12.5	0.3mm
电极斜度(°)	30 ± 1	30.40	30.70	7'

备注:

Notes:

结论(Conclusion): 所校准的项目符合规定的技术指标 The calibration items meet the provisions of the technical indicators.

1.本报告中的扩展不确定度是由标准不确定度乘以包含概率约为95%时的包含因子 k 。

The expanded uncertainty is given in the report by the standard uncertainty multiplied by the probability of about 95% when the factor k .

2.依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

3.第5项电压降项目未获得CNAS认可

(以下空白)

(The below is blank)