



校准证书

CALIBRATION CERTIFICATE



扫一扫验真伪
1GA23111450281-0001

证书编号:
Certificate No.



委托方:
Client

EMC-Testcenter AG

委托方地址:
Address

Moosackerstrasse 77, 8105 Regensdorf, Switzerland

仪器/样品名称:
Description

Test Pin of IEC 61032 Figure8/9

型号/规格:
Model/Type

SMT-1312

制造厂商:
Manufacturer

LISUN GROUP

出厂编号:
Serial No.

LS2311131046

管理号:
Asset No.

/

样品接收日期:
Date of Receipt

2023-11-15
Y M D

结果:
Conclusion

按实测结果使用,见报告内页(Use according to the measured results, See the page of the report)

校准日期:
Date of Calibration

2023-11-15
Y M D

建议下次校准日期:
Due Date

2024-11-14
Y M D

校准:
Calibration by

吴娟

审核:
Inspected by

胡中兴

授权签字人:
Approved Signatory

方文潮



本实验室地址: 广东省广州市番禺区石碁镇农科所南街8号 @力赛计量实验室

Address: No.8.South Street Shi Ji Institute Guangzhou.Guangdong.China

联系电话(Tel): 020-31134076

投诉电话(CT): 020-31104772

传真(Fax): 020-31134076

邮政编码(Post): 511400

公司网址(Web): www.lisaitest.com

电子邮件(E-mail): cal@lisaitest.com



校准说明

DIRECTIONS OF CALIBRATION

证书编号:
Certificate No.



1GA23111450281-0001

本机构质量管理体系符合ISO/IEC17025标准要求。

The quality system is in accordance with ISO/IEC17025.

1. 证书内页中"P"代表"合格","F"代表"不合格","N/A"代表"不适用"。

In the datasheet,"P" represents "Pass" and "F" represents "Fail" and "N/A" represents "Not Applicable".

2. 本证书编号具有唯一性, 后缀若带有"A~Z"的证书为替换证书, 自发出后原证书即刻作废。

Each certificate has a unique number,The suffix of "A~Z" will be added to the number as a replacement of the old version. The original certificate will be officially invalid once the new certificate number is issued.

3. 环境条件。(Environmental condition during the calibration).

温度(Temperature): 20.4°C 相对湿度 (Relative Humidity): 59 %

4. 校准地点。(Place of the Calibration).

广州总部实验室(101精测室)

5. 未加盖“证书报告专用章”无效,报告无校准或检测、审核、授权签字人签章无效。

The report is invalid without the official stamp, report is invalid without 'report stamp', The report is invalid without the signatures of Approval and Reviewer.

6. 样品及生产/研制方的相关信息由委托方提供并由委托方承担相应责任。

The information of the samples and the manufacturer/developer shall be provided by the applicants and applicant shall bear the corresponding responsibilities.

7. 本次校准的技术依据及CNAS认可范围, 超出范围的内容未被认可。详细认可范围请查看CNAS网站中注

册编号为L7127的证书附件。(Reference documents and accredited scope by CNAS for the Calibration beyond which isn't accredited. Please see the attachment of certificate No.L7127 on CNAS website for details.)

JJF (辽) 288-2016 外壳对人和设备的防护检验用试具校准规范 Calibration Specification for Protection of person and equipment by enclosures-Probe for verification

8. 本次校准使用的主要计量标准器具。(Main standards of measurement used in the Calibration).

名称 Description	出厂编号 Serial No.	证书号/有效期 Certificate No./ Due Date	溯源机构 Traceability Institute	技术特征 Technique Character
万能工具显微镜	3875	CYY202300319 / 2024-03-23	华南国家计量测试中心	MPE: ± (1+L/100)μm; L: mm
数显卡尺	K18A305745	3GC23020139650-2404 / 2024-02-05	广州力赛计量检测有限公司	(0~70) mm, MPE=±0.02mm; (> 70~200) mm, MPE=±0.03mm; (> 200~300) mm, MPE=±0.04mm



证书编号:
Certificate No.




1GA23111450281-0001

1、外观以及一般性检查: 正常
In view of External and Generality check : Pass

2、尺寸(Dimension):

2.1、示意图 (Diagram)



2.2、尺寸(Dimension):

标称值 Nominal (mm)	实测值 Measured (mm)	不确定度 Uncertainty (mm)	技术要求 Requirement (mm)	结论 Conclusion (Pass/Fail)
Φ25	25.03	0.04	24.8 ~ 25.2	P
10	9.99	0.04	/	/
Φ3	2.998	0.006	2.95 ~ 3.00	P
Φ4	3.950	0.006	3.95 ~ 4.00	P
20	20.18	0.04	/	/
4	3.93	0.04	/	/
15	14.953	0.006	14.9 ~ 15.0	P

注：技术要求由客户提供

附： 关于测量结果不确定度的说明：

appendix: Directions of uncertainty in the calibration

1.依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

2.本报告给出的扩展不确定度是由合成标准不确定度乘以包含概率约为95%时对应的包含因子k得到的。

The expanded uncertainty given in this report is obtained by multiplying the combined standard uncertainty by the corresponding coverage factor k when the coverage probability of about 95%.

-----以下空白(The below is blank)-----

