



校准证书
CALIBRATION CERTIFICATE



扫一扫验真伪
2GB25072819621-0003
校验码:A250C66A

证书编号:
Certificate No.

委托方:
Client Saudi Standards, Metrology and Quality Organization (SASO)

委托方地址:
Address Riyadh - Al Muhammadiyah, P.O. Box 3437, Riyadh 11471, Kingdom of Saudi Arabia

仪器/样品名称:
Description 水接触灯具试验装置
Test Device for Contact Water Luminaire

型号/规格:
Model/Type LS-B58

制造厂商:
Manufacturer LISUN GROUP

出厂编号:
Serial No. LS2508050957 管理号:
Asset No. /

样品接收日期:
Date of Receipt 2025-08-05
Y M D

结果:
Conclusion 所校准项目合格(Passed at Calibration Items)

校准日期:
Date of Calibration 2025-08-05
Y M D

校准:
Calibration by 郭亲 郭亲

审核:
Inspected by 潘玉全 潘玉全

授权签字人:
Approved Signatory 方文潮 方文潮



本实验室地址: 广东省广州市番禺区石碁镇农科所南街8号

Address: No.8.South Street Shi Ji Institute Guangzhou.Guangdong.China

联系电话(Tel): 020-31134076

投诉电话(CT): 020-39969860

传真(Fax): 020-39969986

邮政编码(Post): 511400

公司网址(Web): www.lisaitest.com

电子邮件(E-mail): cal@lisaitest.com

我司电子证书具备网站/微信公众号查验(二维码扫码查验)、校验码、电子签章(查验内容是否修改且一致)三重防伪验证。



校准说明

DIRECTIONS OF CALIBRATION

证书编号:
Certificate No.



2GB25072819621-0003
校验码:A250C66A

本实验室质量管理体系符合ISO/IEC17025标准要求。

The quality system is in accordance with ISO/IEC17025.

1. 本实验室出具的数据均可溯源至国际单位制(SI)。

All data issued by this laboratory are traceable to International System of Units(SI)

2. 证书中最大允许误差、判定结果仅供参考,其中"P"代表"合格","F"代表"不合格","N/A"代表"不适用"。

使用人员应结合实际测量需求,评估测量不确定度对符合性评定的影响。

MPE & judgement result in the datasheet is only for reference,"P" represents "Pass" and "F" represents "Fail" and "N/A" represents "Not Applicable".Where as users should evaluate the effects of MU of calibration results on conformance assessment by actual measurement.

3. 本证书编号具有唯一性,后缀若带有"A~Z"的证书为替换证书,自发出后原证书即刻作废。

Each certificate has a unique number,The suffix of "A~Z" will be added to the number as a replacement of the old version. The original certificate will be officially invalid once the new certificate number is issued.

4. 校准地点、环境条件。(Place and environmental condition of the Calibration).

客户现场(1102试验室)

温度(Temperature): 23°C 相对湿度 (Relative Humidity): 65 %

5. 未加盖“证书报告专用章”无效,报告无校准或检测、审核、授权签字人签章无效。

The report is invalid without the official stamp, report is invalid without 'report stamp', The report is invalid without the signatures of Approval and Reviewer.

6. 本次校准的技术依据及CNAS认可范围,超出范围的内容未被认可。详细认可范围请查看CNAS网站中注

册编号为L7127的证书附件。(Reference documents and accredited scope by CNAS for the Calibration beyond which isn't accredited. Please see the attachment of certificate No.L7127 on CNAS website for details.)

JJG 617-1996 数字温度指示调节仪检定规程 V. R. of Digital Temperature Indicators and controllers

7. 计量溯源性声明:Measurement traceability declaration:

被校器具 (Description): 水接触灯具试验装置(Test Device for Contact Water Luminaire); 被校器具

(Description) → 高精度多功能过程校验仪 → 多功能校准仪(万用表部分)(Multi-Product Calibrator) → 华南国家计量测试中心(South China National Center of Metrology)

8. 本次校准使用的主要计量标准器具。(Main standards of measurement used in the Calibration).

名称/型号 Equipment/Model	出厂编号 Serial No.	证书号/有效期 Certificate No./ Due Date	溯源机构 Traceability Institute	技术特征 Technique Character
高精度多功能过程校验仪/726	7678586	3NC25040703234-6 944 / 2026-05-06	广州力赛计量检测有限公司 (Guangzhou LiSai Metrology & Test Co., Ltd.)	DCV: $U=0.004\%$; DCV: $U=0.001\%$; OHM: $U=0.01\%$; ACV: $U=0.01\%$; ACV: $U=0.09\%$; FEQ: $1 \times 10^{-7} (k=2)$ 温 度: $U=0.3^{\circ}\text{C}$, $k=2$



广州力赛计量检测有限公司

Guangzhou LiSai Metrology & Test Co., Ltd.



2GB25072819621-0003

校验码:A250C66A

证书编号:

Certificate No.



1、外观以及一般性检查: 正常

In view of External and Generality check: Pass

2、温度示值误差 (Temperature indication error) :

标准值	指示值	误差	U	允许误差	结论
Reference	Indicated	Error	($k=2$)	MPE	Conclusion
(°C)	(°C)	(°C)	(°C)	(°C)	(P/F)
15.0	15.2	0.2	0.3	± 2.0	P
20.0	20.2	0.2	0.3	± 2.0	P
25.0	24.9	-0.1	0.3	± 2.0	P
30.0	29.7	-0.3	0.3	± 2.0	P
35.0	34.4	-0.6	0.3	± 2.0	P

注: 校准线型为K型。

Note: The calibration line type is K-type.

按客户要求点校准。

Calibrate as per the customer's requirements.

附: 关于测量结果不确定度的说明:

appendix: Directions of uncertainty in the calibration

1.依据(Reference document)

JJF 1059.1-2012测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

2.本报告给出的扩展不确定度是由合成标准不确定度乘以包含概率约为95%时对应的包含因子k得到的。

The expanded uncertainty given in this report is obtained by multiplying the combined standard uncertainty by the corresponding coverage factor k when the coverage probability of about 95%.

-----以下空白(The below is blank)-----